




REVISIONS				
ZONE	REV	DESCRIPTION	DATE	APPROVED

11G03N004 A



- Notes:
- 1) Material: Epoxy, Polyurethane, & PVC
 - 2) For best accuracy, Suspended Solids measurements should be taken 2ft. below the surface and 3ft. from the wall.



UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES. TOLERANCES ON: FRACTIONS ANGLES		CONTRACT NO.		 80 Whisperwood Blvd. Suite 107 Slidell, LA 70458	
MATERIAL		DRAWN BY		DATE	
SPEC		CHECKER		DESIGNER	
FINISH		PROJ. ENGR.		RELEASED	
		APPROVAL			
TSS SENSOR OUTLINE & MOUNTING		SIZE	FSCM NO.	DRAWING NO.	REV
		B		11G03N004	A
SCALE		NTS		SHEET 1 OF 1	